



How to Search for a Certificate of Analysis



If your internet browser is
“Microsoft Internet Explorer”

1. Use the search function in your internet browser for the lot number, or
2. Download the pdf and then use the search function in your pdf reader for the lot number.



If your internet browser is
“Mozilla Firefox” or other

1. Download the pdf file in Acrobat Reader, and
2. Use the search or find function in Acrobat Reader for the lot number



Lot Number MD1722A004

Certificate of Analysis

Product name: **ActeroTM MediaboxTM Salmonella**
 Product number: FCM-047
 Lot number: MD1722A004
 Expiry date: 2018-06-20
 Storage conditions: 2-25°C (36-77°F)


Physical Characteristics

<u>Specification</u>	<u>Expected results</u>	<u>Results</u>
Prepared: Appearance	Clear with or without a light precipitate	Clear
Color	Light to medium yellow	Light yellow
pH:	8.2 ± 0.2	8.2

Microbiological performance

<u>Microorganisms</u>	<u>Approx. Inoculum</u> (CFU)	<u>Expected Results</u>	<u>Results</u>
<i>S. Typhimurium</i> ATCC 14028	10-100	Good growth	Good growth
<i>E. coli</i> O157:H7 ATCC 43895	10-100	Good growth	Good growth

Laboratory Results: Satisfactory



Quality Assurance/Quality Control

2017-10-06

Date

