



Lot Number MD1822A003

**Quality Control Document**

Actero Salmonella Enrichment  
Batch Number: 161128011912  
FSI Catalog Number: FCM-047  
Lot Number: MD1822A003  
Expiration Date: 13SEP2019

**Physical and Chemical Tests**

Appearance:  
Color:  
pH at 25°C:

Transparent Liquid  
Light Amber  
8.2 +/-0.2

**Microbiological Tests**

Microorganisms

*Escherichia coli* ATCC 25922  
*E. coli* O157:H7 ATCC 35150  
*Salmonella poona* NCTC 4840

Growth\*

Excellent  
Excellent  
Very Good

\*The following results were obtained from type cultures after incubation at a temperature of 35°C and observed after 18-24 hours.

Sterility (5 days at 37°C)

Clear (Not cloudy)

**Laboratory Results: Satisfactory**

Country of Manufacture:  
Testing performed by:  
Completed on:  
QA Release by:

USA  
Gina Moon  
December 17<sup>th</sup> 2018  
Stuart Clark

