

CERTIFICATE OF ANALYSIS

PRODUCT DATA	
Product	ACTERO LISTERIA
Code	FCM-204
Batch	103755
Presentation	2 Prepared Bags /5L/Bags 5L
Expiry date	31/01/2020
Manufacturing date	05/02/2019

FORMULA*	in g / L
Formula according to customer specifications	
*Adjusted and/or supplemented as required to meet performance criteria	

PHYSICAL-CHEMICAL TEST	
Colour	Intense Yellow
pH	7,2 (7,2 ± 0,2) 5000 (5000 ± 15)
Result: Satisfactory	

STERILITY TEST
Incubation 48 hours at 30-35°C and 48 hours at 20-25°C: NO GROWTH Check at 7 days after incubation in same conditions
Result: Satisfactory

PERFORMANCE TEST				
Prepare Tubes - Inoculate with 100±20 CFU for Growth Promotion or 10 ⁴ -10 ⁶ CFU for Selectivity Microbiological control according to ISO 11133:2014/ Adm 1:2018.				
Aerobiosis. Incubation at 30 ± 1 °C during 18-24 h				
Microorganism	Specification	Inoculated CFU	Results	
<i>Listeria monocytogenes</i> ATCC® 13932, WDCM 00021 + (1) + (2)	> 10 CFU. Blue-green coln. w. opaque halo (Ottaviani Agosti)	89	Conforms	
<i>Listeria monocytogenes</i> ATCC® 35152, WDCM 00109 + (1) + (2)	> 10 CFU. Blue-green coln. w. opaque halo (Ottaviani Agosti)	88	Conforms	
<i>Enterococcus faecalis</i> ATCC® 19433, WDCM 00009	Inhibited	10000	Conforms	
<i>Escherichia coli</i> ATCC® 8739, WDCM 00012	Inhibited	10000	Conforms	
				Result: Satisfactory

This certificate does not release the user from the reception control upon receipt of the good.

This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.

Quality release date: 08/02/2019