



Lot Number MD1928B009

Quality Control Document
Actero Listeria Enrichment
Batch Number: 180303012103
Expiry: 03/2021
Lot Number: MD1928B009
Expiration Date: 02FEB2020

Physical and Chemical Tests

| | |
|-------------|--------------------|
| Appearance: | Transparent Liquid |
| Color: | Light Amber |
| pH at 25°C: | 7.15 |

Microbiological Tests

| | |
|---|----------------|
| <u>Microorganisms</u> | <u>Growth*</u> |
| <i>Escherichia coli</i> 0157:H7ATCC 35150 | No Growth |
| <i>L. innocua</i> ACTC 33090 | Excellent |

*The following results were obtained from type cultures after incubation at a temperature of 35°C and observed after 18-24 hours.

| | |
|----------------------------|--------------------|
| Sterility (5 days at 37°C) | Clear (Not cloudy) |
|----------------------------|--------------------|

Laboratory Results: Satisfactory

| | |
|-------------------------|---------------------------|
| Country of Manufacture: | USA |
| Testing performed by: | Gina Moon |
| Completed on: | MAY 06 th 2019 |
| QA Release by: | Stuart Clark |

