



Lot Number MD1928A002

Quality Control Document
Actero Listeria Enrichment
Batch Number: 180302012103
Expiry: 03/2021
Lot Number: MD1928A002
Expiration Date: 22NOV2019

Physical and Chemical Tests

Appearance:	Transparent Liquid
Color:	Light Amber
pH at 25°C:	7.14

Microbiological Tests

<u>Microorganisms</u>	<u>Growth*</u>
<i>Escherichia coli</i> 0157:H7ATCC 35150	No Growth
<i>L. innocua</i> ACTC 33090	Excellent

*The following results were obtained from type cultures after incubation at a temperature of 35°C and observed after 18-24 hours.

Sterility (5 days at 37°C)	Clear (Not cloudy)
----------------------------	--------------------

Laboratory Results: Satisfactory

Country of Manufacture:	USA
Testing performed by:	Gina Moon
Completed on:	FEB. 26 th 2019
QA Release by:	Stuart Clark

