



Lot Number MD1911B002

**Quality Control Document**

Sterile Water  
Catalog number 67-C0071  
FSI Catalog Number: FCM-032  
Lot Number: MD1911B002  
Expiration Date: 12FEB2020

**Water Treatment**

Processing (1):	Passed through a mixed-bed Ion Exchange Resin
Effect of Processing (1):	Removal of mineral ions present in the water
Processing (2):	Sterile Filtration (0.22uM)
Effect of Processing (2):	Removes any bacterial

**Physical and Chemical Tests**

Appearance:	Transparent Liquid
Color:	Clear

**Sterility Tests**

<u>Incubation</u>	<u>Growth</u>
Sample incubated for 72 hours at 37°C	No Growth
Media Box incubated for 3 days at Room Temperature	No Growth

**Laboratory Results: Satisfactory**

Country of Manufacture:	USA
Testing performed by:	Stuart Clark
Completed on:	FEB 14 <sup>th</sup> 2019
QA Release by:	Stuart Clark

