



Certificate of Analysis

Actero Salmonella Supplement 1

Lot Number: ME28

Expiry: Nov 8, 2019

Physical and Chemical Tests:

<u>Properties</u>	<u>Specification</u>	<u>Analysis</u>
Appearance:	Clear, colorless viscous Liquid	Pass
Color:	≤15R	3
Monoethanolamine:	≥99.5%	99.6
Diethanolamine	≤0.2%	0.2%
Specific Gravity:	1.016 – 0.019	1.016-1.019
Water Content:	≤0.3%	0.2%

Sterility Tests:

<u>Microorganisms / Sterility</u>	<u>Growth</u>
Liquid Sterility Test (48 hours at 36°C ± 1°C)	Clear / no cloudiness
Aerobic count (48 hours at 36°C ± 1°C on TSA plate)	No Colonies

Laboratory Results: Satisfactory

Country of Manufacture: USA
Testing performed by: Stuart Clark
Completed on: Nov 14, 2018
QA Release by: Stuart Clark