



Lot Number MD1922A005

**Quality Control Document**

Actero Salmonella Enrichment  
Batch Number: 180723042108  
FSI Catalog Number: FCM-047  
Lot Number: MD1922A005  
Expiration Date: 15AUG2020

**Physical and Chemical Tests**

|             |                    |
|-------------|--------------------|
| Appearance: | Transparent Liquid |
| Color:      | Light Amber        |
| pH at 25°C: | 8.2 +/- 0.2        |

**Microbiological Tests**

Microorganisms

*Escherichia coli* ATCC 25922  
*E. coli* O157:H7 ATCC 35150  
*Salmonella poona* NCTC 4840

Growth\*

Excellent  
Excellent  
Very Good

\*The following results were obtained from type cultures after incubation at a temperature of 35°C and observed after 18-24 hours.

|                            |                    |
|----------------------------|--------------------|
| Sterility (5 days at 37°C) | Clear (Not cloudy) |
|----------------------------|--------------------|

**Laboratory Results: Satisfactory**

|                         |                           |
|-------------------------|---------------------------|
| Country of Manufacture: | USA                       |
| Testing performed by:   | Gina Moon                 |
| Completed on:           | NOV 19 <sup>th</sup> 2019 |
| QA Release by:          | Stuart Clark              |

