



Lot Number MD2028B012

**Quality Control Document**  
Actero Listeria Enrichment  
Batch Number: 200307012303  
Expiry: 03/2023  
Lot Number: MD2028B012  
Expiration Date: 01MAR2021

### Physical and Chemical Tests

Appearance:	Transparent Liquid
Color:	Light Amber
pH at 25°C:	7.09

### Microbiological Tests

<u>Microorganisms</u> <i>Escherichia coli</i> 0157:H7ATCC 35150 <i>L. innocua</i> ACTC 33090	<u>Growth*</u> No Growth Excellent
--	--

\*The following results were obtained from type cultures after incubation at a temperature of 35°C and observed after 18-24 hours.

Sterility (5 days at 37°C)	Clear (Not cloudy)
----------------------------	--------------------

### Laboratory Results: Satisfactory

Country of Manufacture:	USA
Testing performed by:	Gina Moon
Completed on:	June 03 <sup>rd</sup> 2020
QA Release by:	Stuart Clark

