

CERTIFICATE OF ANALYSIS

PRODUCT DATA

Product	ACTERO LISTERIA
Code	FCM-204
Batch	121559
Presentation	2 Prepared Bags /5L/Bags 5L
Expiry date	19/02/2022
Manufacturing date	24/02/2021

FORMULA*

in g / L

Formula according to customer specifications

* Adjusted and/or supplemented as required to meet performance criteria

PHYSICAL-CHEMICAL TEST

Colour	Intense Yellow
pH	7,3 (7,2 ± 0,2)
	5000 (5000 ± 15)

Result: **Satisfactory**

STERILITY TEST

Incubation 48 hours at 30-35 °C and 48 hours at 20-25 °C: NO GROWTH.

Check at 7 days after incubation in same conditions.

Result: **Satisfactory**

PERFORMANCE TEST

Prepare Tubes - Inoculate with 100±20 CFU for Growth Promotion or 10⁴-10⁶ CFU (selectivity).

Microbiological control according to ISO 11133:2014/A1:2018.

Aerobiosis. Incubation at 30 ± 1 °C during 24 ± 2 h.

Microorganism	Specification	Inoculated CFU	Results
<i>Listeria monocytogenes</i> ATCC® 13932, WDCM 00021 + (1) + (2)	> 10 CFU. Blue-green coln. w. opaque halo (Ottaviani Agosti)	58	Complies
<i>Listeria monocytogenes</i> ATCC® 35152, WDCM 00109 + (1) + (2)	> 10 CFU. Blue-green coln. w. opaque halo (Ottaviani Agosti)	89	Complies
<i>Enterococcus faecalis</i> ATCC® 19433, WDCM 00009	Inhibited	53000	Complies
<i>Escherichia coli</i> ATCC® 8739, WDCM 00012	Inhibited	98000	Complies

Result: **Satisfactory**

This certificate does not release the user from the reception control upon receipt of the good. This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.

Quality release date: 02/03/2021

CERTIFICADO DE ANÁLISIS

DATOS DE PRODUCTO

Producto	ACTERO LISTERIA
Código	FCM-204
Lote	121559
Presentación	2 Bolsas /5L/Bolsa 5L
Caducidad	19/02/2022
Fabricación	24/02/2021

FORMULA*

en g / L

Fórmula según especificaciones del cliente

CONTROL FÍSICO-QUÍMICO

Color	Amarillo limón
pH	7,3 (7,2 ± 0,2)
	5000 (5000 ± 15)

Resultado: **Satisfactorio**

*Ajustado y/o suplementado según los criterios de recuperación necesarios

CONTROL DE ESTERILIDAD

Incubación 48 horas a 30-35 °C y 48 horas a 20-25 °C: SIN CRECIMIENTO

Verificación a 7 días tras incubación en las mismas condiciones.

Resultado: **Satisfactorio**

CONTROL DE RENDIMIENTO

Dosificar Tubos - Inocular 100±20 UFC para Productividad o 10⁴-10⁶ UFC para Selectividad.

Control microbiológico según normativa ISO 11133:2014/ A1:2018.

Aerobiosis. Incubación a 30 ± 1 °C. Lectura a 24 ± 2 h.

Microorganismo	Especificación	UFC inoculados	Result.
<i>Listeria monocytogenes</i> ATCC® 13932, WDCM 00021 + (1) + (2)	> 10 UFC. Coln. azul-verdoso. Halo opaco (A. Ottaviani	58	Conforme
<i>Listeria monocytogenes</i> ATCC® 35152, WDCM 00109 + (1) + (2)	> 10 UFC. Coln. azul-verdoso. Halo opaco (A. Ottaviani	89	Conforme
<i>Enterococcus faecalis</i> ATCC® 19433, WDCM 00009	Inhibido	53000	Conforme
<i>Escherichia coli</i> ATCC® 8739, WDCM 00012	Inhibido	98000	Conforme

Resultado: **Satisfactorio**

Este certificado no exime al usuario de realizar el control de recepción al recibir el producto

Este certificado es una copia electrónica del certificado disponible en nuestro laboratorio y no requiere firma.

Fecha: 02/03/2021