



Lot Number MD2128B044

**Quality Control Document**

Actero Listeria Enrichment  
Batch Number: 210521022405  
FSI Catalog number: FCM-046  
Expiry: 05/2024  
Lot Number: MD2128B044  
Expiration Date: 22APR2022

**Physical and Chemical Tests**

Appearance:  
Color:  
pH at 25°C:

Transparent Liquid  
Light Amber  
7.11

**Microbiological Tests**

Microorganisms

*Escherichia coli* 0157:H7ATCC 35150  
*L. innocua* ACTC 33090

Growth\*

No Growth  
Excellent

\*The following results were obtained  
from type cultures after incubation at  
a temperature of 35°C and observed  
after 18-24 hours.

Sterility (5 days at 37°C)

Clear (Not cloudy)

**Laboratory Results: Satisfactory**

Country of Manufacture:  
Testing performed by:  
Completed on:  
QA Release by:

USA  
Gina Moon  
JUL 26<sup>th</sup> 2021  
Stuart Clark

