

Quality control certificate



Actero Listeria

Nº Cat: 2074

Nº Batch: 401251

Re-test Date: 1/2027

Formula in g/l

Typical formula g/L * Adjusted and/or supplemented as required to meet performance criteria.

Quality control

Solubility	Appearance	Color of the dehydrated medium	Color of the prepared medium	Final pH (25°)	Batch pH
w/o rests	Fine powder	Beige	Amber	7,0±0,2	7,2

Microbiological test

According ISO 11133:

Incubation conditions: Target microorganisms (37 °C /18 h) / Non-target microorganisms (37 °C/24 h).

Condiciones de inoculación: Target microorganisms (<100 CFU) / Non-target microorganisms (> 10000 CFU) / Selectividad (> 10000 CFU).

Microorganisms	Specification
Enterococcus faecalis ATCC 19433	<100 colonies on TSA
Escherichia coli ATCC 25922	<10 colonies on TSA
Listeria monocytogenes ATCC 13932 + Escherichia coli ATCC 25922 + Enterococcus faecalis ATCC 19433	>100 colonies on ALOA
Listeria monocytogenes ATCC 35152 + Escherichia coli ATCC 25922 + Enterococcus faecalis ATCC 19433	>100 colonias on ALOA
Listeria grayi ATCC 25401 + Escherichia coli ATCC 25922 + Enterococcus faecalis ATCC 19433	>100 colonias on ALOA
Listeria ivanovii ATCC 19119 + Escherichia coli ATCC 25922 + Enterococcus faecalis ATCC 19433	>100 colonies on ALOA

This certificate is an electronic copy of the certificate available in our laboratory and does not require signature.